


MIPS Calibration at 24 and 70 μ m

Alberto Noriega-Crespo (MIPS Lead)

on behalf of the Instrument Team (Arizona) &
Instrument Support Team (Spitzer Science Center)



PUBLICATIONS OF THE ASTRONOMICAL SOCIETY OF THE PACIFIC, 119: 994–1018, 2007 September
© 2007, The Astronomical Society of the Pacific. All rights reserved. Printed in U.S.A.

Absolute Calibration and Characterization of the Multiband Imaging Photometer for *Spitzer*. I. The Stellar Calibrator Sample and the 24 μ m Calibration

C. W. ENGELBRACHT,¹ M. BLAYLOCK,¹ K. Y. L. SU,¹ J. RHO,² G. H. RIEKE,¹ J. MUZEROLLE,¹ D. L. PADGETT,² D. C. HINES,²
K. D. GORDON,¹ D. FADDA,² A. NORIEGA-CRESPO,¹ D. M. KELLY,¹ W. B. LATTER,³ J. L. HINZ,¹ K. A. MISSEL,¹
J. E. MORRISON,¹ J. A. STANSBERRY,¹ D. L. SHUPP,² S. STOLOVY,² WM. A. WHEATON,² E. T. YOUNG,¹ G. NEUGEBAUER,¹
S. WACHTER,² P. G. PÉREZ-GONZÁLEZ,^{1,4} D. T. FRAYER,² AND E. R. MARLEAU²

PUBLICATIONS OF THE ASTRONOMICAL SOCIETY OF THE PACIFIC, 119: 1019–1037, 2007 September
© 2007, The Astronomical Society of the Pacific. All rights reserved. Printed in U.S.A.

Absolute Calibration and Characterization of the Multiband Imaging Photometer for *Spitzer*. II. 70 μ m Imaging

KARL D. GORDON,¹ CHARLES W. ENGELBRACHT,¹ DARIO FADDA,² JOHN STANSBERRY,¹ STELANIE WACHTER,² DAVE T. FRAYER,²
GEORGE RIEKE,¹ ALBERTO NORIEGA-CRESPO,² WILLIAM B. LATTER,³ ERICK YOUNG,¹ GERRY NEUGEBAUER,¹ ZOLTAN BALOG,¹
JEFFREY W. BEEMAN,⁴ HÉRYÉ DOLE,¹ EICHI EGAMI,¹ EUGENE E. HALLER,^{4,5} DEAN HINZ,² DOUG KFLY,¹
FRANÇOISE MARLEAU,² KARL MISSEL,¹ JANE MORRISON,¹ PABLO PÉREZ-GONZÁLEZ,^{1,4} JEONGHIL RHO,²
AND WM. A. WHEATON²

Outline

- MIPS Summary
- Zero Points (Rieke Photometric System)
- Repeatability
- Linearity
- Comparison with IRAS
- Extended Source Calibration

MIPS Summary

MULTIBAND IMAGING PHOTOMETER FOR SPITZER (MIPS)
Pocket Guide

<http://ssc.spitzer.caltech.edu/mips/>

Basic MIPS Capabilities:

Imaging photometry at 24, 70, and 160 μm and low resolution ($R = 15 - 25$) spectroscopy between 52 and 97 μm . A fine pixel scale option at 70 μm (no change required for the other two bands), combined with precise subpixel sampling dither patterns, allow data processing to achieve "super resolution" imaging. A cryogenic scan mirror mechanism provides freeze frame scan mapping, efficient dithering, and other instrument capabilities.

MIPS Instantaneous Fields of View:

24 μm	5.4x5.4 arcminutes
70 μm	5.25x2.6 or 2.6x1.3 arcminutes
160 μm	0.53x5.33 arcminutes (effective)
SED Slit	2.0x0.33 arcminutes (full wavelength coverage)

Basic Sensitivities (low background):
 5 sigma in 500 seconds on source

24 μm	110 μJy
70 μm default	7.2 mJy
70 μm fine	14.4 mJy
SED	82/201/447 mJy @ 60/75/90 μm
160 μm	29 (40 w/ confusion) mJy

The MIPS Astronomical Observation Templates:

- Photometry & Super Resolution**
 - Telescope staring mode imaging photometry
- Scan Mapping**
 - Freeze frame mapping in all three bands with constant telescope slewing
- Spectral Energy Distribution (SED)**
 - Low resolution ($R = 15 - 25$) spectroscopy over 52 to 97 μm
- Total Power Mode**
 - Zero level brightness of very extended emission

The MIPS Detector Arrays:

24 μm	SiAs (IBC) 128x128 pixels; 2.55" 4.7 μm bandwidth
70 μm	Ge Ga 32x32 pixels; 5.3" or 9.96" 19 μm bandwidth SED $R = 15 - 25$ (9.8" pixels)
160 μm	Stressed Ge Ga 2x20 pixels; 16.0"x18.0" 35 μm bandwidth

Saturation Limits:

Point source in 1 second (Jy); Extended source in 10 seconds (MJy/ster)

Band	Point Source	Extended
24 μm	4.1	260
70 μm (default)	23	101
70 μm (fine)	57	292
SED @ 60,75,90 μm	250/290/1000	1087/1261/4350
160 μm	3	20



Prepared by:
 William B. Latter
 SSC/MIPS Instrument Support Team Lead
 (v7.0, updated 31 October 06, JWC)



24 μ m Calibration - Summary

- The primary calibrators are A stars (22 sample)
- Calibration factor (from instrumental units to astrophysical units) = $4.54 \cdot 10^{-2}$ MJy/sr/DN/s
- Absolute uncertainty = 4%
- After 2.5yrs: 348 measurements of 141 stars (covering a factor 460 in flux density) from 8.6mJy up to 4 Jy
- Repeatability for HD 159330 (routine calibrator) shows a rms scatter of 0.4%
- The calibration is based on the standard 3sec photometric observation (agrees within 1% with Scan mode)

24 μ m Post-Processing




Fig. 1.— Artifacts not discussed by Gordon et al. (2005) that are now fixed as part of standard processing with the MIPS Data Analysis Tool (DAT), illustrated using observations of HD 159330 (AOR key 13587712) plotted in reverse gray scale. (a) The jailbar effect is most easily seen in individual frames, here caused by a cosmic ray below and to the right of the star, and (b) fixed as described in § 2. (c) The spots caused by debris on the pick-off mirror (several sharp white spots above and to the right of the star, as well as diffuse white regions below and to the left of the star) are most easily seen in a mosaicked image and are (d) fixed using separate flat fields for each scan mirror position.






TABLE 3
DATA USED TO COMPUTE THE 24 μm CALIBRATION FACTOR

Name	K_s^* (mag)	24 μm Count Rate (DN s $^{-1}$)	Uncertainty (DN s $^{-1}$)	Calibration Factor (MJy sr $^{-1}$ [DN s $^{-1}$] $^{-1}$)	Uncertainty (MJy sr $^{-1}$ [DN s $^{-1}$] $^{-1}$)
HD 000319	5.479	6.708E+03	5.00E+01	4.513E-02	1.39E-03
HD 002811	7.057	1.591E+03	3.35E+01	4.448E-02	1.63E-03
HD 011413	5.422	7.866E+03	3.87E+01	4.056E-02	1.23E-03
HD 014943	5.439	7.180E+03	2.84E+01	4.374E-02	1.32E-03
HD 015646	6.411	2.997E+03	4.85E+01	4.281E-02	1.46E-03
HD 017254	5.877	4.529E+03	1.97E+01	4.632E-02	1.40E-03
HD 020888	5.691	5.519E+03	2.04E+01	4.512E-02	1.36E-03
HD 021981	5.526	6.089E+03	1.33E+01	4.761E-02	1.43E-03
HD 034868	6.024	3.851E+03	3.89E+01	4.758E-02	1.51E-03
HD 042525	5.751	5.116E+03	2.85E+01	4.606E-02	1.40E-03
HD 057336	7.114	1.403E+03	3.44E+01	4.786E-02	1.86E-03
HD 073210	6.165	3.707E+03	5.55E+01	4.341E-02	1.45E-03
HD 073666	6.532	2.541E+03	5.66E+01	4.517E-02	1.69E-03
HD 073819	6.280	3.170E+03	5.29E+01	4.566E-02	1.57E-03
HD 092845	5.513	7.166E+03	4.51E+01	4.094E-02	1.26E-03
HD 101452	6.819	1.910E+03	3.50E+01	4.613E-02	1.62E-03
HD 105805	5.600	5.734E+03	5.10E+01	4.722E-02	1.48E-03
HD 116706	5.502	6.188E+03	5.01E+01	4.789E-02	1.49E-03
HD 128998	5.756	4.824E+03	2.79E+01	4.862E-02	1.48E-03
HD 158485	6.145	3.564E+03	2.79E+01	4.599E-02	1.43E-03
HD 163466	6.339	2.917E+03	2.94E+01	4.699E-02	1.49E-03
HD 172728	5.753	4.747E+03	2.21E+01	4.954E-02	1.50E-03


NOTE.—The average calibration factor is 4.54×10^{-2} MJy sr $^{-1}$ (DN s $^{-1}$) $^{-1}$, to which we have assigned an uncertainty of 2% (see § 3).
* $K_s = [24]$ for the stars in this table (see G. H. Rieke et al. 2007, in preparation), all of which are between types A0 and A6.

The zero point is derived from a set of well measured A type stars (22)

Repeatability is 0.4% over the first 1000 days.



Departure from linearity is 0.3% between 9 and 4000 mJy



IRAS 12 & 25 μ m
and MIPS 24 μ m

[point sources]

Measured:


$$F(24\mu\text{m})/F(12\mu\text{m}) = 0.265$$

$$F(24\mu\text{m})/F(25\mu\text{m}) = 1.11$$


Predicted (Kurucz's Models):

$$F(24\mu\text{m})/F(12\mu\text{m}) = 0.266$$


$$F(24\mu\text{m})/F(25\mu\text{m}) = 1.11$$



No obvious effects due to background, but the measurement is noisy.



24 μ m Extended Source Calibration



75 SINGS galaxies plus M31 & LMC
MIPS/IRAS = 0.96 (within 8% instrumental uncertainties)

Fig. 10.— Ratio of MIPS to IRAS measurements of extended sources as a function of flux density measured at 24 μ m. The error bars represent the combined uncertainty on both measurements. The dashed line is drawn at 1 as a guide.

70um Calibration - Summary

- The predicted colors at 70 μ m are derived from those 24 μ m (ratio of the 23.675 and 71.42 μ m effective wavelengths with the appropriate Kurucz model).
- The Wide Field (WF) calibration factor is 702 +/- 35 Mjy/sr/MIPS70 based on 66 stars.
- Absolute uncertainty = 7%
- The WF calibration covers a flux density range 22mJy to 17Jy
- The zero point at 71.42 μ m (Rieke Photometric System) is 0.778+/- 0.012
- The photometric repeatability is 4.5%
- The calibration is based on the standard WF photometry observations (3 and 10sec)

70 μ m Post-Processing




Fig. 1.— Mosaics of coarse-scale observations of two stars of different brightness shown with (a, d) default reductions, (b, e) column mean subtraction, and (c, f) column mean subtraction and time filtering. The two stars are HD 197989 (AOR key 13590784, $\mathit{flux} \times \mathit{density} = 787$ mJy) and HD 53501 (AOR key 13641984, $\mathit{flux} \times \mathit{density} = 135$ mJy). The HD 197989 images are displayed with a linear stretch that ranges (a) from 0.022 to 0.05, (b) from -0.005 to 0.025, and (c) from -0.005 to 0.0025. The HD 53501 images are displayed with a linear stretch that ranges (d) from 0.009 to 0.025, (e) from -0.005 to 0.01, and (f) from -0.002 to 0.008. The images and ranges are all given in MIPS70 units.

